

## VWR® LABORATORY BOTTLES, HIGH-DENSITY POLYETHYLENE, WIDE MOUTH

As per the manufacturer, the below product meets the following criteria:

<b>VWR Catalog Number</b>	414004-113
<b>VWR Europe Catalog Number</b>	215-3440
<b>Lot Number</b>	VHF180618
<b>Description</b>	VWR BOTTLE HDPE WM 250ML PK12
<b>Date of Manufacture</b>	June 18, 2018
<b>Country of Origin</b>	India

### Quality System Compliance

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2008 & ISO 13485:2003 Quality Management System. Inspection records are reviewed and signed for product release.

### QC Testing

Representative production samples are collected and inspected in accordance with current applicable standard protocols.

Test Name	Test Reference	Accuracy	Result
<b>Leak Test</b>	QWI-QC-03	General Insp. Level - I, AQL - 1.5	Leak proof
<b>Autoclaving</b>	QWI-QC-07	General Insp. Level - I, AQL - 1.5	NA
<b>Dimensional</b>	QSF-QC-02	General Insp. Level - I, AQL - 1.5	Within tolerance

### Product Specifications

**Material:** **Bottle:** High-density Polyethylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.

**Closure:** Polypropylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.

**BSE/TSE Statement:** The above-mentioned product is manufactured without the use of Raw materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of the manufacturer's knowledge.

**Non-Pyrogenic:** Not Applicable.

**DNase & RNase Free:** Not Applicable.

**Cytotoxicity:** Not Applicable.

**Latex-Free Statement:** No materials containing latex or natural rubber are used in the manufacturing, handling and packaging processes for this product.

Signed:



Jamie Ethier  
VP Global Quality  
VWR, Part of Avantor

**Date:** June 18, 2018