

Corning Incorporated - Life Sciences  
 Antigua Carretera A Roma KM 3.6  
 San Nicolas De Los Garza NL 66480 MEX  
[www.corning.com/lifesciences](http://www.corning.com/lifesciences)  
 Refer to website for regional contact information.

Page: 1 / 1

|                        |                                |                                      |
|------------------------|--------------------------------|--------------------------------------|
| <b>Product Name</b>    | : 9825 RG SCT 25X150 TUBE COMP | W/ SCREW CAP                         |
| <b>Catalog Number</b>  | : 9825-25                      | <b>Manufacture Date</b> : 2018-09-10 |
| <b>Lot ID</b>          | : 25318028                     |                                      |
| <b>Expiration Date</b> | : Unconfirmed Validity         |                                      |

**Quality Management System** - Complies with the current version of the ISO 9001 Standard.

**Glass Type** - Borosilicate, Type I, Class A, ASTM E438 - Standard Specification for Glasses in Laboratory Apparatus.

**Quality Control Testing** - Representative production samples are collected and inspected in accordance with current applicable product specifications. Inspection records are reviewed and approved by qualified personnel for product release. Key inspections and inline tests are listed below:

Visual Inspection - Pass  
 Packaging Inspection - Pass  
 Dimensional Test - Pass

- This product met Corning Incorporated - Life Sciences' high standards of quality at the time of batch/lot release.



Miguel Flores  
 Quality Manager