

Date of Certificate: 10/31/2018
Product Number: 5000-0020
Product Lot Number: 1239682
Specified Dose Range: 19.0 - 28.0 KGY

Run ID	Receipt Date	Quantity	UOM	Delivered Min	Delivered Max
2313667	10/30/2018	156	CS	20.0	26.3
2312787	10/30/2018	156	CS	20.3	25.9



Barry Srolis
Sr. Quality Manager