

## VWR® LABORATORY BOTTLES, HIGH-DENSITY POLYETHYLENE

This document certifies that the below product meets the following criteria:

<b>VWR Catalog Number</b>	414004-157
<b>VWR Europe Catalog Number:</b>	215-3406
<b>Lot Number</b>	VDV111018
<b>Description</b>	VWR Bottle Hdpe Nm 250ml Pk12.
<b>Date of Manufacture:</b>	October 11, 2018
<b>Country of Origin:</b>	India.

### Quality System Compliance

Products are manufactured, inspected and found to be in compliance with product and quality specification requirements as documented in the ISO 9001:2015 & ISO13485:2016 Quality Management System. Inspection records are reviewed and signed for product release.

### QC Testing

Representative production samples are collected and inspected in accordance with current applicable standard protocols.

Test Name	Test Reference	Accuracy	Result
<b>Leak Test</b>	QWI-QC-03	General Insp. Level- I, AQL- 1.5	Leak proof
<b>Autoclaving</b>	QWI-QC-07	General Insp. Level- I, AQL- 1.5	NA
<b>Dimensional</b>	QSF-QC-02A	General Insp. Level- I, AQL- 1.5	Within tolerance

### Product Specifications

<b>Material:</b>	Bottle- High-Density Polyethylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed. Closure- Polypropylene, 100 % Virgin, FDA 21 CFR Compliance, USP Class VI passed.
<b>BSE/TSE Statement:</b>	The above-mentioned product is manufactured without the use of Raw materials of animal origin. Therefore, BSE/TSE issue does not concern this product to the best of the manufacturer's knowledge.
<b>Non-Pyrogenic:</b>	Not Applicable.
<b>DNase &amp; RNase Free:</b>	Not Applicable.
<b>Cytotoxicity:</b>	Not Applicable.
<b>Latex-Free Statement:</b>	No materials containing latex or natural rubber are used in the manufacturing, handling and packaging processes for this product.

Signed:



Jamie Ethier  
VP Global Quality  
VWR, Part of Avantor

**Date:** December 7, 2018