

VWR® SPECTROSIL SPECTROPHOTOMETER CELLS

As per the manufacturer, the below product meets the following criteria:

| | |
|---------------------|-------------------------------|
| VWR Catalog Number | 414004-062 |
| Lot Number | 2019011321 |
| Description | VWR CELL QUARTZ FLAT 10MM PK2 |
| Date of Manufacture | January 13, 2019 |
| Country of Origin | Made in China |

Quality System Compliance

Products are manufactured under the ISO 9001:2008 standard. Products are Inspected and controlled through whole production processing in accordance with current applicable product specifications and QC SOP. Inspection records are reviewed and signed off by qualified personnel for product release.

QC Testing

Representative production samples are collected and inspected in accordance with current applicable product specifications.

Product Specifications

| | |
|--------------------------|---|
| Path Length of Interior: | 10 +/- 0.05 mm |
| Path Width of Interior: | 10 +/- 0.20 mm |
| Path Length of Exterior: | 12.5 +/- 0.10 mm |
| Path Width of Exterior: | 12.5 +/- 0.10 mm |
| Height of Exterior: | 45 +/-0.5 mm |
| Transmission (at 200nm): | >80% |
| Material: | Quartz |
| BSE/TSE Statement: | The above product contains no ingredients of animal origin and no material derived from, or exposed to animals affected by or under quarantine for Transmitting, Animal Spongiform Encephalopathy (TSE)/Bovine Spongiform Encephalopathy (BSE). |
| Non-Pyrogenic: | N/A |
| DNase & RNase Free: | N/A |
| Cytotoxicity: | N/A |
| Latex Free Statement: | N/A |

Signed:



Jamie Ethier
VP Global Quality
VWR, Part of Avantor

Date: February 8, 2019