



Quality Certificate

VWR® SIGNATURE™ ERGONOMIC HIGH-PERFORMANCE SINGLE-CHANNEL VARIABLE VOLUME PIPETTERS

As per the manufacturer, this product meets the following criteria:

North American Catalog No:	89079-964
European Catalog No:	613-5260
Serial Number:	742733456
Description:	VWR Pipette Ehp Variabl 2-20ul

The enclosed liquid handling instrument has been manufactured in a facility that employs a registered and certified quality management system. This system ensures that the development and production of this product has been strictly controlled and that it meets the highest quality standards. Each instrument passes stringent validation and Control Procedures in accordance with EN ISO 8655.

Volume (µl):	2	20
Accuracy (%):	-1.6766	0.0123
Precision (%):	1.0191	0.1715
Status:	Passed	Passed

Calibrated and tested using the gravimetric method with distilled water:

Tested by: C10

Condition of measurements:

Basis of adjustment: Ex; Reference temperature: 20 °C; Relative air humidity: 50%; Barometric pressure: 101kPa

Signed:

Jamie Ethier
VP Global Quality
VWR, Part of Avantor

Date: March 18, 2019